

----- ACQUISITION PARAMETERS -----

File Name = 1d_spectrum.590
Author =
Sample ID = ugi1.4tms
Content = Single Pulse Experiment
Creation Date = 5-NOV-2009 07:11:13

Revision Date = 2-DEC-2009 00:13:58
Spec Site = Eclipse+ 300

Spec Type = DELTA_NMR
Data Format = 1D COMPLEX
Dimensions = X
Dim Title = 1H
Dim Size = 16384
Dim Units = [ppm]
Acq_delay = 0.2195[ms]
Changer_sample = 0
Experiment = single_pulse.exp
Field_strength = 7.0586013[T]
Irr90 = 9.4[us]
Irr90_hi = 17[us]
Irr90_lo = 46[us]
Irr_pwidth = 46[us]
Lock_status = IDLE
Recvr_gain = 14
Relaxation_delay = 4[s]
Scans = 128
Solvent = CHLOROFORM-D
Spin_get = 16[Hz]
Spin_lock_90 = 1[us]
Spin_lock_attn = 29[dB]
Spin_set = 15[Hz]
Spin_state = SPIN ON
Spin_status = SPIN ON
Temp_get = 22.6[dC]
Temp_set = 25[dC]
Temp_state = TEMP OFF
Temp_status = TEMP OFF
X90 = 9.4[us]
X90_hi = 17[us]
X90_lo = 46[us]
X_acq_duration = 3.6339712[s]
X_domain = 1H
X_freq = 300.52965592[MHz]
X_offset = 5[ppm]
X_points = 16384
X_prescans = 0
X_pulse = 4.7[us]
X_resolution = 0.27518105[Hz]
X_sweep = 4.50856628[kHz]

